



CERTIFICATE

No. Z1 12 08 67069 002

Holder of Certificate: Linet spol. s.r.o.

Želevcice 5 274 01 Slaný CZECH REPUBLIC

Production Facility(ies):

67069

Certification Mark:



Product:

Medical Beds

Application environment:

Model(s):

Eleganza 1

Parameters:

Safe working load:

Max. patient weight:

Dimensions:

Mattress support platform:

Height adjustment:

Rated voltage:

Rated frequency:

250 kg

185 kg

185 kg

218 x 100 cm

200 x 90 cm

40 and 76 cm

230 VAC

830 VAC

840 Fated frequency:

50/60 Hz

1/2/3

max. 370 VA

10%, max 2min/18min

IPX4

Rated power: Degree of protection: Protection class:

Applied part: Type B

Mode of operation:

Angle A: 73°
Angle B: 18,5°
Angle C: 90°
Angle D: 23°

Tested according to:

EN 60601-2-52:2010 EN 60601-1:2006

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.:

713010512

Date, 2012-08-23

(Torsten Zimmer)

Page 1 of 1